

ISO 9001:2015 Co.



Rehsiff Scientific®

Manufacture & Importer of
Laboratory Fine Chemicals, Diagnostic Kit & Lab wares

CERTIFICATE OF ANALYSIS

ISO 9001-2015 CERTIFIED

Product Name : ALUMINIUM AAS STANDARD SOLUTION 1000 mg/L Al IN DILUTED HCl traceable to NIST
NIST Expiry date: 12 months from the date of opening

Batch No.	: Sample COA	Release Date	: 18-Mar-2025
Mol. Formula	:	Mol. Weight	:
Code No.	: RSH002S	CAS No.	: NA
Mfg. Date	: Mar-2025	Exp. Date	: Feb-2029
HAZ. / P.G.	: 8 / III	UN No.	: 3264

Sr.	Tests	Specifications	Results
1	Appearance	A clear colourless liquid	A clear colourless liquid
2	Amount of substance concentration c(Al)	1000 mg/L +/-0.2%	999.38mg/L
3	Matrix	2% HCl	Complies

CONCLUSION -This above product complies as per the specifications of Rehsiff Scientific.

